

**TECHNICAL SPECIFICATIONS (ITB 20 - 0040 - price schedule no. 1)**

| Item No. | Item & Description  | Specification | Statement of Compliance  |
|----------|---|---------------|--|
|          |   |               | <p>Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1 (a)(ii).</p> |
| 1        | <p><b>AMBULANCE</b>, Vehicle unit specifications<br/>                     Brand New<br/>                     Body type: Van<br/>                     Transmission: Automatic<br/>                     Transmission category: Minimum of 4 speed; Maximum of 6-speed<br/>                     Displacement: Minimum of 2,500cc - Maximum of 3,000cc</p> <p>Dimension:<br/>                     Length: Minimum of 5,380mm - Maximum of 5,915mm<br/>                     Width: Minimum of 1,880mm - Maximum of 1,950mm<br/>                     Height: Minimum of 2,280mm - Maximum of 2,285mm<br/>                     Wheelbase: Minimum of 3,110mm - Maximum of 3,860mm</p> <p>Ground clearance: 185 mm to 195mm</p> <p>Engine type, 4 cylinder, 16 valve, Diesel<br/>                     Color: White<br/>                     Emission Standard: EURO IV<br/>                     Transmission: Dash Mounted Automatic<br/>                     Fuel System: Direct Injected Diesel<br/>                     Fuel tank capacity: at least 70 liters<br/>                     Suspension: Front: Coil Spring with Stabilizer bar<br/>                     Rear: Rigid axle, leaf spring<br/>                     Brakes: Front: Ventilated discs, Rear: Leading-Trailing Drum-<br/>                     Steering System: rack and Pinion with Power Assist<br/>                     Audio System: 2 DIN audio, CD/MP3/AUX/USB with remote control<br/>                     SRS Airbag: Driver and Passenger<br/>                     Ambulance Equipment: with top light slim type<br/>                     With Medical Cabinet</p> <p>COMPLYING DOH REQUIREMENT- For Specialty Level 3 and Level 2 Hospitals<br/>                     -Type 11 Ambulance- Advance Life Support (ALS) Ambulance</p> <p>B- INTERIOR SPECIFICATION:</p> |               |  |

**TECHNICAL SPECIFICATIONS (ITB 20 - 0040 - price schedule no. 1)**

| Item No. | Item & Description  | Specification | Statement of Compliance |
|----------|---|---------------|-------------------------|
|          | <p>-Partition sealed- separating driver &amp; patient, mounted and secured</p> <p>- Full body wall medicine cabinet with "pure sine" power inverter 300W</p> <p>-IV hook mounted</p> <p>- Patient chart holder mounted</p> <p>- 3 seats for Nurses w/ back rest for comfort</p> <p>-overhead hand grab rail ceiling mounted (stainless steel)</p> <p>-(3) garbage bin holder mounted (green, yellow and black)</p> <p><b>MEDICAL EQUIPMENT &amp; SUPPLIES</b></p> <p><b>A- TRANSPORT EQUIPMENT</b></p> <p>1- Stretcher 2 link type, Sub stretcher pole type</p> <p>1 Ortho scope stretcher</p> <p>1 spinal board</p> <p><b>B-VENTILATION, AIRWAY AND APPARATUS</b></p> <p>1 Laryngoscope set pedia and adult</p> <p>1 Portable Suction apparatus with accessories, compatible to newborn, pedia and adult</p> <p>1 Portable oxygen tank 20 lbs w/ accessories and content</p> <p>1 Portable oxygen tank, 5 lbs w/ complete accessories (oxygen gauge)</p> <p>1 Pedia Resuscitator</p> <p>1 Adult Resuscitator</p> <p>1 Heavy duty nebulizer, branded</p> <p><b>C- MONITORING &amp; DEFIBRILLATOR APPATUS</b></p> <p>1 Manual Defibrillator with cardiac monitor with disposable pads, pedia &amp; adult</p> <p>1 Blood pressure apparatus wall mounted with Pedia and adult cap with extender cap for obese and large arms</p> <p>1 Pulse Oximeter compatible for adult and pedia</p> <p>1 Stethoscope, compatible for pedia and adult (branded)</p> <p><b>D: OBSTETRICAL DELIVERY SET</b></p> <p>1 delivery bag/kit with sterile content</p> <p>1 set baby wrap, blanket, cap , mittens for newborn</p> <p><b>E: MISCELLANEOUS SUPPLIES</b></p> <p>1- Weighing scale, heavy duty, bathroom</p> <p>1- Thermometer, non-mercurial</p> <p>1- Heavy bandage paramedic scissors for cutting of clothes, belts &amp; boots</p> <p><b>ACCESSORIES &amp; COMMUNICATION EQUIPMENT</b></p> <p>With flashing light and Siren with Public Address System</p> <p>2- units handheld radio with NTC License</p> <p>1 unit Mobile phone with car charger</p> <p>1- Dash cam (front and rear view)</p> <p>2- pcs Heavy duty flashlight</p> <p>1- 5 lbs fire extinguisher (rating 2A10BC)</p> <p>1 pc Portable chiller for vaccine</p> <p>Ambulance wheeled cot with mounted cot fastening system</p> <p>Intra-vehicle intercom system (between driver and the</p> |               |                         |

**TECHNICAL SPECIFICATIONS (ITB 20 - 0040 - price schedule no. 1)**

| Item No. | Item & Description   | Specification | Statement of Compliance |
|----------|--|---------------|-------------------------|
|          | <p>body)</p> <p>INCLUSIONS:<br/>                     -Ambulance sticker and signage (left, right, front and rear).<br/>                     -Logo of end -user (prescribed DOH logo on both side of the vehicle).<br/>                     - GSIS Comprehensive Insurance &amp; LTO registration in the name of "Provincial Government of South Cotabato".<br/>                     - to be delivered at the Provincial Capitol Compound, Koronadal City, South Cotabato.</p> <p>Note: Bidder Must:<br/>                     1- Specify the brand and model of equipment to be installed.<br/>                     2- Bidder must specify the type of materials to used for the installation such as in Flooring, cabinet, partition, etc.<br/>                     3- Delivery period- 30 days upon receipt of Purchase Order</p> <p>Terms &amp; Condition:<br/>                     Provide Training for manual operation and trouble shooting<br/>                     At least 1 year warranty on parts – indicated in the after sales warranty certificate<br/>                     At least 3 years warranty on services – indicated in the after sales warranty certificate<br/>                     Minimum of 5 years for spare parts availability – indicated in the after sales warranty certificate</p> <p>Preferably with service center or sub-office at General Santos City, Davao City and nearby cities in Mindanao.</p> <p>If supplier is a distributor, all warranties best offered by the principal must also be offered by the distributor<br/>                     Verified List of Facilities where offered products were installed for post qualification purposes and evaluation.</p> |               |                         |

*x-x-x nothing follows x-x-x*

\_\_\_\_\_  
**Name and Signature of Bidder**